

# **MLA0320 Animal Nerve Stimulating Electrode**

## Accessories

The MLA0320 Animal Nerve Stimulating Electrode consists of two exposed gold-plated brass electrodes. The rod has a plastic handle for easy handling and precise positioning. The electrodes are separated by 3 mm (0.1") and rounded at the tip. The device can be connected to the BNC Output of any PowerLab. It is ideal for experiments that involve nerve and muscle stimulation in animals such as rodents and amphibians. NOT FOR USE ON HUMANS.



### Caution

Read "Statement of Intended Use" on our website or in "Getting Started with PowerLab" before use.

#### **Ordering Information:**

MLA0320 Animal Nerve Stimulating Electrode

For use with: Powerlab

Specification: Gold-plated brass electrodes Length of cable: 1.2 m (4') Length of handle: 200 mm (7.9")

Electrode length:10 mm (0.4") Electrode separation: 3 mm (0.1") Electrode diameter: 2 mm (0.08")

#### **ADINSTRUMENTS.com**

#### ISO 9001:2008 Certified Quality Management System

WARRANTY: ADInstruments PowerLab data acquisition units (Product Number Prefix: PL), Front-end Signal Conditioners (Product Number Prefix: FE) are warranted against defects in materials and workmanship for a period of 5 years from the date of purchase. Other PowerLab data acquisition units and ADInstruments manufactured Front-end and Pod Signal Conditioners, and Instruments are warranted of a period of 3 years from the date of purchase. Third party products are covered by the manufacturer's warranty. Warranties are void if the product has been damaged due to negligence. Consumables and electrodes are not covered by a warranty. All questions regarding service and warranty should be directed to your nearest ADInstruments representative.