

MLT118F GSR Finger Electrodes

Transducer Series

Bipolar finger electrodes for measuring galvanic skin response. Suitable for connection to the FE116 GSR Amp. MLT118F is MR unsafe, see MLT117F for use in MR environments.



Caution

Read “Statement of Intended Use” on our website or in “Getting Started with PowerLab” before use.

Ordering Information:

MLT118F GSR Finger Electrodes

For use with:

FE116 GSR Amplifier