

# **MLAWBT9 EEG Flat Electrodes**

## Accessories

The MLAWBT9 EEG Flat Electrodes are a set of five gold plated, cupshaped disk recording electrodes suitable for recording EEG, EMG and ECG. The electrodes have a diameter of 9 mm (0.35"). They feature a hole in the top for injection of conductive paste or gel and lead wires of

185 cm (6.1'), fitted with shrouded 1.5 mm (0.06") sockets for connection to MLA2340 or MLA2540 Shielded Bio Amp Cables. The MLAWBT9 EEG Flat Electrodes can also be used with the GT205 8 Channel, GT201 16 Channel, FE238 Octal and FE234 Quad Bio Amps.

## Caution

Read "Statement of Intended Use" on our website before use. Leads are not supplied sterile. Appropriate Infection Control Procedures should be followed between subsequent subject connections. **Clean with water and mild detergent. Do not use alcohol.** 

### **Ordering Information:**

MLAWBT9 EEG Flat Electrodes

For use with: MLA2340 3 Lead Shielded Bio Amp Cable MLA2540 5 Lead Shielded Bio Amp Cable GT205 8 Channel Bio Amp GT201 16 Channel Bio Amp FE238 Octal Bio Amp





NOTE: Lead Wires are connected to the top row of signal pins in Shielded Bio Amp Cables

### ADINSTRUMENTS.com

ISO 9001:2008 Certified Quality Management System

WARRANTY: ADInstruments PowerLab data acquisition units (Product Number Prefix: PL), Front-end Signal Conditioners (Product Number Prefix: FE) are warranted against defects in materials and workmanship for a period of 5 years from the date of purchase. Other PowerLab data acquisition units and ADInstruments manufactured Front-end and Pod Signal Conditioners, and Instruments are warranted of a period of 3 years from the date of purchase. Third party products are covered by the manufacturer's warranty. Warranties are void if the product has been damaged due to negligence. Consumables and electrodes are not covered by a warranty. All questions regarding service and warranty should be directed to your nearest ADInstruments representative.

